



प.सं./Ref. No.:

मिति/Date : 12/23/2025

बिषय: बोलपत्र संशोधन तथा म्याद थप सम्बन्धी सूचना ।

यस क्याम्पस अन्तर्गत तपसिल बमोजिमको मालसामान खरिद कार्यको लागि बोलपत्र आह्वान गरिएकोमा, सो बोलपत्र सम्बन्धी कागजातहरु (Technical Specifications & Price Schedule) तथा बोलपत्र दाखिला गर्ने अन्तिम मिति देहाय बमोजिम संशोधन गरिएको व्यहोरा सम्बन्धित सबैको जानकारीको लागि यो सूचना प्रकाशित गरिएको छ ।

✓ म्याद थप सम्बन्धी विवरण

सार्वजनिक खरिद ऐन तथा नियमावली अनुसार बोलपत्र पेश गर्ने अन्तिम मिति र समय देहाय बमोजिम कायम गरिएको छ:

ठेक्का नं.	खरिद कार्यको नाम	बोलपत्र दाखिला गर्ने अन्तिम मिति र समय	बोलपत्र खोल्ने मिति र समय
GAC/G/NCB-01/2082-83	SUPPLY, DELIVERY AND INSTALLATION OF LABORATORY EQUIPMENT AND MODULAR LABORATORY SETUP	2025-12-28 AD, 12:00 hours (२०८२ पुष १३ गते)	2025-12-28 AD, 13:00 hours (२०८२ पुष १३ गते)

✓ प्राविधिक संशोधन

Pursuant to the ITB (Instructions to Bidders) regarding the amendment of Bidding Documents, this Addendum is issued to modify, clarify, and amend the following aspects of the Tender Documents. This Addendum shall form an integral part of the Bidding Documents.

1. General Amendments & Clarifications for civil works

S.N.	Reference Section	Existing Clause / Context	Modified Clause / Clarification
1	Furniture	Quantity discrepancies for the "Modular Laboratory System"	The Price Schedule of Goods is amended to 12 units, consistent with the Technical Specifications. The

सुबोध कर्मचारी




प.सं./Ref. No.:

मिति/Date : 12/23/2025

1. General Amendments & Clarifications for civil works

S.N.	Reference Section	Existing Clause / Context	Modified Clause / Clarification
1	Furniture	(Price Schedule: 4 vs. Technical Specs: 12 vs. Drawings: 20).	quantity shown in the drawings is for general layout reference only. Bidders must strictly quote for 12 units .
2	Drawings	Drawings were provided in the original bidding document for the modular laboratory setup.	Drawings Withdrawn: The drawings provided in the original bid document were for administrative and illustrative purposes only and are hereby withdrawn . Bidders are requested to rely strictly on the technical specifications and Schedule of Requirements.
3	Plumbing Works	"Plumbing Works" not being listed as separate items in the Price Schedule/BOQ.	Plumbing works are not listed as separate line items. All detailed plumbing works described in the Technical Specifications are considered an integral part of the Furniture with Sink. Bidders must include the cost of these plumbing works within the quoted rates for the corresponding furniture items.
4	Site Visit	No specific provision for site inspection for furniture and AC installation.	Bidders are encouraged to visit the laboratory site at Gauradaha Agriculture Campus to inspect the location for the "Installation of modular laboratory system" and "Air Conditioner installation" to ensure accurate measurements and compatibility before submitting their bids.
5	UPS Backup	UPS 10KVA	The Price Schedule includes separate line items for "Air Conditioners" and "UPS". Bidders must quote separately for 4 Units of Air Conditioners and 1 Unit of Central UPS (10kVA) as specified.


सुबोध खनाल
क्याम्पस प्रमुख



प.सं./Ref. No.:

मिति/Date : 12/23/2025


2. Amendments to Technical Specifications for Laboratory Equipments

Based on queries regarding technical discrepancies, the following specifications are amended:

S.N.	Item Name	Original Specification / Issue	Amended Specification
1	Kjeldahl Apparatus	Specification mentioned only "Distillation set" was ambiguous regarding the full system	Distillation + Digestion + Scrubber: The specification is amended to include the complete system: Digestion Unit, Distillation Unit, and Scrubber.
2	Autoclave	Horizontal electrically heated autoclave is required.	Horizontal / Vertical: The specification is amended to read: Microprocessor/PLC controlled Horizontal or Vertical electrically heated autoclave is required.
3	Trinocular Stereo Zoom Microscope	Included ambiguous text: Variable Intensity control dial approx 24999 hours and 3.3watt-30-degree oblique illumination	Deleted: The specific text regarding "24999 hours..." is removed to eliminate ambiguity.
4	Homogenizer	Included text: Count Capacity: upto 6 digits X 1.7 Surface Magnification	Deleted: This text appears to be a clerical error (referring to a colony counter) and is hereby deleted from the Homogenizer specification.
5	Digital Fruit Penetrometer	Discrepancy between "Digital" in the name and "Analog" in the display type specification.	Digital: The Display Type is clarified to be Digital, consistent with the item name.

Note to Bidders:

1. All other terms and conditions of the Bidding Documents remain unchanged.
2. Bidders are required to acknowledge the receipt of this Addendum and include it in their Bid Response.



सुबोध खनाल

क्याम्पस प्रमुख

सुबोध खनाल

क्याम्पस प्रमुख

Ph. No.: +977-23-480275 Email: info@gac.tu.edu.np Website: https://gac.tu.edu.np